

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090924\
 Data File : VV037233.D
 Acq On : 09 Sep 2024 10:58
 Operator : SY/MD
 Sample : P3843-06
 Misc : 5.00mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 09 22:03:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Sep 09 22:02:35 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	599481	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	577642	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	267895	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	263537	48.057	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	96.120%	
7) Chloroethane-d5	1.532	69	210536	49.676	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	99.360%	
11) 1,1-Dichloroethene-d2	2.056	65	111239	48.263	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	96.520%	
21) 2-Butanone-d5	3.789	46	179122	84.348	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	84.350%	
24) Chloroform-d	4.240	84	441378	51.427	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.860%	
26) 1,2-Dichloroethane-d4	4.931	65	275795	51.027	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.060%	
32) Benzene-d6	4.950	84	890797	51.057	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.120%	
36) 1,2-Dichloropropane-d6	5.982	67	287128	52.543	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	105.080%	
41) Toluene-d8	7.236	98	777291	49.364	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	98.720%	
43) trans-1,3-Dichloroprop...	7.548	79	124567	49.271	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	98.540%	
47) 2-Hexanone-d5	8.021	63	114338	76.665	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	76.660%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	298827	45.088	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	90.180%	
66) 1,2-Dichlorobenzene-d4	11.554	152	283934	52.919	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	105.840%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090924\
Data File : VV037233.D
Acq On : 09 Sep 2024 10:58
Operator : SY/MD
Sample : P3843-06
Misc : 5.00mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 09 22:03:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Sep 09 22:02:35 2024
Response via : Initial Calibration

