

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091224\
 Data File : VV037255.D
 Acq On : 12 Sep 2024 09:23
 Operator : SY/MD
 Sample : VV0912MBL01
 Misc : 5.0G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK204

Quant Time: Sep 13 00:43:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 13 00:43:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	660338	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.777	117	639447	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	305293	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	247516	40.976	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.960%
7) Chloroethane-d5	1.532	69	202997	43.483	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.960%
11) 1,1-Dichloroethene-d2	2.057	65	108400	42.697	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.400%
21) 2-Butanone-d5	3.790	46	174224	74.481	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	74.480%
24) Chloroform-d	4.240	84	439267	46.464	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.920%
26) 1,2-Dichloroethane-d4	4.934	65	266905	44.831	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.660%
32) Benzene-d6	4.950	84	888350	45.996	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.000%
36) 1,2-Dichloropropane-d6	5.982	67	287611	47.544	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.080%
41) Toluene-d8	7.236	98	771843	44.280	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.560%
43) trans-1,3-Dichloroprop...	7.548	79	122399	43.734	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.460%
47) 2-Hexanone-d5	8.024	63	103959	62.968	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	62.970%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	293214	39.965	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	79.940%
66) 1,2-Dichlorobenzene-d4	11.555	152	288355	47.160	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091224\
Data File : VV037255.D
Acq On : 12 Sep 2024 09:23
Operator : SY/MD
Sample : VV0912MBL01
Misc : 5.0G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK204

Quant Time: Sep 13 00:43:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 13 00:43:12 2024
Response via : Initial Calibration

