

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091324\
 Data File : VV037274.D
 Acq On : 13 Sep 2024 10:14
 Operator : SY/MD
 Sample : VSTD00116
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001216

Quant Time: Sep 14 02:24:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 14 02:23:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	15894	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	15145	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	7886	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	17017	1.001	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.945	65	9897	1.021	ug/L	-0.02
7) 1,2-Dichloropropane-d6	5.995	67	12938	1.029	ug/L	0.00
8) Toluene-d8	7.238	98	37005	1.003	ug/L	-0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	9480	1.007	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	15233	0.986	ug/L	99
6) Trichloroethene	5.828	95	11632	0.983	ug/L	89
9) 1,2-Dibromoethane	8.290	107	7292	1.018	ug/L #	98
12) 1,2,3-Trichloropropane	10.197	75	6734	1.039	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.356	75	1123	0.983	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091324\
Data File : VV037274.D
Acq On : 13 Sep 2024 10:14
Operator : SY/MD
Sample : VSTD00116
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001216

Quant Time: Sep 14 02:24:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 14 02:23:22 2024
Response via : Initial Calibration

