

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091324\
 Data File : VV037279.D
 Acq On : 13 Sep 2024 14:13
 Operator : SY/MD
 Sample : P3990-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AS0

Quant Time: Sep 14 03:50:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Sep 14 03:38:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	15611	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	15157	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	7700	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	7955	0.476	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	96.000%		
4) 1,2-Dichloroethane-d4	4.969	65	4618	0.485	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	98.000%		
7) 1,2-Dichloropropane-d6	5.996	67	5857	0.466	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	94.000%		
8) Toluene-d8	7.254	98	16078	0.435	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	88.000%		
10) 1,1,2,2-Tetrachloroeth...	10.146	84	4227	0.449	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	90.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV091324\
Data File : VV037279.D
Acq On : 13 Sep 2024 14:13
Operator : SY/MD
Sample : P3990-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AS0

Quant Time: Sep 14 03:50:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM091324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Sep 14 03:38:32 2024
Response via : Initial Calibration

