

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091624\
 Data File : VV037298.D
 Acq On : 16 Sep 2024 11:25
 Operator : SY/MD
 Sample : P3934-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 17 05:06:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 17 05:05:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	303884	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	297764	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	137863	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	228911	82.348	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 164.700%#		
7) Chloroethane-d5	1.532	69	187321	87.191	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 174.380%#		
11) 1,1-Dichloroethene-d2	2.060	65	98786	84.551	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 169.100%#		
21) 2-Butanone-d5	3.796	46	150403	139.717	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 139.720%#		
24) Chloroform-d	4.243	84	386829	88.913	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 177.820%#		
26) 1,2-Dichloroethane-d4	4.937	65	236895	86.464	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 172.920%#		
32) Benzene-d6	4.953	84	796113	88.520	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 177.040%#		
36) 1,2-Dichloropropane-d6	5.982	67	253896	90.132	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 180.260%#		
41) Toluene-d8	7.239	98	688978	84.883	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 169.760%#		
43) trans-1,3-Dichloroprop...	7.551	79	105085	80.634	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 161.260%#		
47) 2-Hexanone-d5	8.024	63	90118	117.220	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 117.220%		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	262491	76.833	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 153.660%#		
66) 1,2-Dichlorobenzene-d4	11.554	152	245760	89.007	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 178.020%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091624\
Data File : VV037298.D
Acq On : 16 Sep 2024 11:25
Operator : SY/MD
Sample : P3934-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 17 05:06:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 17 05:05:17 2024
Response via : Initial Calibration

