

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091624\
 Data File : VV037294.D
 Acq On : 16 Sep 2024 09:21
 Operator : SY/MD
 Sample : VV0916MBL01
 Misc : 5.0G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK206

Quant Time: Sep 17 05:05:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 17 05:05:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	695560	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	659516	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	305621	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	297601	46.773	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.540%
7) Chloroethane-d5	1.532	69	233585	47.501	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.056	65	123148	46.049	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.100%
21) 2-Butanone-d5	3.789	46	187890	76.255	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	76.260%
24) Chloroform-d	4.236	84	482068	48.409	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.820%
26) 1,2-Dichloroethane-d4	4.931	65	294825	47.013	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.020%
32) Benzene-d6	4.950	84	984358	49.416	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.840%
36) 1,2-Dichloropropane-d6	5.979	67	316547	50.735	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.480%
41) Toluene-d8	7.236	98	865863	48.163	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.320%
43) trans-1,3-Dichloroprop...	7.548	79	133715	46.324	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.640%
47) 2-Hexanone-d5	8.021	63	113175	66.464	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	66.460%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	318980	42.154	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.300%
66) 1,2-Dichlorobenzene-d4	11.554	152	312593	51.069	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091624\
Data File : VV037294.D
Acq On : 16 Sep 2024 09:21
Operator : SY/MD
Sample : VV0916MBL01
Misc : 5.0G/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK206

Quant Time: Sep 17 05:05:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 17 05:05:17 2024
Response via : Initial Calibration

