

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV091719\
 Data File : VV012885.D
 Acq On : 18 Sep 2019 01:30
 Operator : SY/MD
 Sample : K4834-09
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Sep 18 02:31:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 18 01:09:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	371619	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	375373	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	141387	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	193871	5.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.40%
7) Chloroethane-d5	1.58	69	153509	5.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.80%
11) 1,1-Dichloroethene-d2	2.13	63	239553	4.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.40%
20) 2-Butanone-d5	3.98	46	403358	52.37	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.74%
24) Chloroform-d	4.40	84	305426	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
26) 1,2-Dichloroethane-d4	5.08	65	163504	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.40%
32) Benzene-d6	5.09	84	630929	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	6.12	67	196689	5.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.80%
41) Toluene-d8	7.36	98	513344	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
43) trans-1,3-Dichloropropene-	7.66	79	61131	4.49	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.80%
46) 2-Hexanone-d5	8.14	63	284720	49.74	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.48%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	143893	4.91	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	176061	5.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV091719\
Data File : VV012885.D
Acq On : 18 Sep 2019 01:30
Operator : SY/MD
Sample : K4834-09
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Sep 18 02:31:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR091719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 18 01:09:35 2019
Response via : Initial Calibration

