

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091824\
 Data File : VV037350.D
 Acq On : 18 Sep 2024 22:42
 Operator : SY/MD
 Sample : P3988-08 50X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH8R7

Quant Time: Sep 19 03:59:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 19 02:54:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	311183	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	316906	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	144775	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	105975	5.805	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	116.000%
7) Chloroethane-d5	1.529	69	87306	4.587	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.800%
11) 1,1-Dichloroethene-d2	2.053	65	45569	4.833	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.600%
20) 2-Butanone-d5	3.812	46	230872	42.315	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.640%
24) Chloroform-d	4.243	84	203304	4.587	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
26) 1,2-Dichloroethane-d4	4.941	65	98869	4.764	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	4.953	84	394320	4.641	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	5.982	67	127260	4.816	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.400%
41) Toluene-d8	7.239	98	334366	4.367	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.561	79	35140	4.043	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.800%
46) 2-Hexanone-d5	8.024	63	196716	46.323	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.640%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	101739	4.724	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.400%
66) 1,2-Dichlorobenzene-d4	11.558	152	128609	4.997	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091824\
Data File : VV037350.D
Acq On : 18 Sep 2024 22:42
Operator : SY/MD
Sample : P3988-08 50X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH8R7

Quant Time: Sep 19 03:59:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 19 02:54:38 2024
Response via : Initial Calibration

