

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091824\
 Data File : VV037334.D
 Acq On : 18 Sep 2024 15:20
 Operator : SY/MD
 Sample : P3988-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH8R2

Quant Time: Sep 19 01:23:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 19 01:19:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	294207	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	310946	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	143755	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	107126	6.206	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	124.200%
7) Chloroethane-d5	1.529	69	87062	4.838	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.800%
11) 1,1-Dichloroethene-d2	2.053	65	46985	5.271	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	105.400%
20) 2-Butanone-d5	3.825	46	220589	42.763	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.520%
24) Chloroform-d	4.243	84	193865	4.626	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	4.940	65	94776	4.830	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
32) Benzene-d6	4.953	84	387202	4.644	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	5.985	67	124263	4.793	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.800%
41) Toluene-d8	7.239	98	331332	4.411	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
43) trans-1,3-Dichloroprop...	7.561	79	39857	4.673	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.400%
46) 2-Hexanone-d5	8.027	63	190345	45.682	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.360%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	97383	4.609	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	11.557	152	127056	4.971	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.847	95	7195	0.285	ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091824\
Data File : VV037334.D
Acq On : 18 Sep 2024 15:20
Operator : SY/MD
Sample : P3988-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH8R2

Quant Time: Sep 19 01:23:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 19 01:19:52 2024
Response via : Initial Calibration





#34

Trichloroethene

Concen: 0.285 ug/L

RT: 5.847 min Scan# 14

Delta R.T. 0.023 min

Lab File: VV037334.D


Acq: 18 Sep 2024 15:20

Instrument :

MSVOA_V

ClientSampleId :

BH8R2



Tgt Ion: 95 Resp: 7195

Ion Ratio Lower Upper

95 100

97 58.0 46.5 86.5

132 94.1 65.5 121.7

130 90.5 71.3 132.3

