

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091824\
 Data File : VV037351.D
 Acq On : 18 Sep 2024 23:06
 Operator : SY/MD
 Sample : P3988-12 50X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH8S1

Quant Time: Sep 19 03:59:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 19 02:54:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	281389	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	291470	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	128468	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	101352	6.139	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.800%
7) Chloroethane-d5	1.529	69	86610	5.032	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.600%
11) 1,1-Dichloroethene-d2	2.053	65	42604	4.997	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.000%
20) 2-Butanone-d5	3.809	46	231547	46.933	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.860%
24) Chloroform-d	4.243	84	198655	4.956	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	4.940	65	96852	5.161	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	4.953	84	384865	4.925	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%
36) 1,2-Dichloropropane-d6	5.985	67	122362	5.035	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	7.239	98	327216	4.647	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
43) trans-1,3-Dichloroprop...	7.561	79	29979	3.750	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.000%
46) 2-Hexanone-d5	8.024	63	198330	50.779	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.560%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	101358	5.117	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.400%
66) 1,2-Dichlorobenzene-d4	11.558	152	120973	5.297	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.000%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.063	96	6001	0.304	ug/L	# 1
19) 1,1-Dichloroethane	3.114	63	5052	0.128	ug/L	95
22) cis-1,2-Dichloroethene	3.815	96	8518	0.380	ug/L	94
34) Trichloroethene	5.831	95	80571	3.405	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV091824\
Data File : VV037351.D
Acq On : 18 Sep 2024 23:06
Operator : SY/MD
Sample : P3988-12 50X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH8S1

Quant Time: Sep 19 03:59:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 19 02:54:38 2024
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 0.304 ug/L

RT: 2.063 min Scan# 319

Delta R.T. -0.003 min

Lab File: VV037351.D

Acq: 18 Sep 2024 23:06

Instrument :

MSVOA_V

ClientSampleId :

BH8S1





Tgt Ion: 96 Resp: 6001

Ion Ratio Lower Upper

96 100

61 357.3 119.9 222.7#

63 1895.2 81.4 151.2#



#19

1,1-Dichloroethane

Concen: 0.128 ug/L

RT: 3.114 min Scan# 645

Delta R.T. 0.006 min

Lab File: VV037351.D

Acq: 18 Sep 2024 23:06




Tgt Ion: 63 Resp: 5052


Ion Ratio Lower Upper

63 100

65 29.8 22.5 41.7

83 15.7 8.9 16.5









#22
cis-1,2-Dichloroethene
Concen: 0.380 ug/L
RT: 3.815 min Scan# 861
Delta R.T. 0.006 min
Lab File: VV037351.D
Acq: 18 Sep 2024 23:06

Instrument :
MSVOA_V
ClientSampleId :
BH8S1

Tgt Ion: 96 Resp: 8518
Ion Ratio Lower Upper
96 100
61 131.8 87.6 162.6
98 60.0 44.9 83.5



#34
Trichloroethene
Concen: 3.405 ug/L
RT: 5.831 min Scan# 1490
Delta R.T. 0.003 min
Lab File: VV037351.D
Acq: 18 Sep 2024 23:06

Tgt Ion: 95 Resp: 80571
Ion Ratio Lower Upper
95 100
97 60.9 46.5 86.5
132 98.2 65.5 121.7
130 99.3 71.3 132.3

