

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092024\
 Data File : VV037440.D
 Acq On : 20 Sep 2024 18:15
 Operator : SY/MD
 Sample : P4081-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AP7

Quant Time: Sep 20 23:39:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 20 23:32:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	294574	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	303403	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	135549	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	57037	3.300	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.000%
7) Chloroethane-d5	1.529	69	75694	4.201	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.000%
11) 1,1-Dichloroethene-d2	2.053	65	32263	3.615	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.200%
20) 2-Butanone-d5	3.796	46	268924	52.069	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.140%
24) Chloroform-d	4.243	84	201252	4.796	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
26) 1,2-Dichloroethane-d4	4.937	65	96669	4.920	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%
32) Benzene-d6	4.953	84	366319	4.503	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
36) 1,2-Dichloropropane-d6	5.986	67	125067	4.944	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.800%
41) Toluene-d8	7.240	98	299403	4.085	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.600%
43) trans-1,3-Dichloroprop...	7.564	79	27147	3.262	ug/L	0.01
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.200%
46) 2-Hexanone-d5	8.024	63	212430	52.250	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.500%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	107612	5.219	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.400%
66) 1,2-Dichlorobenzene-d4	11.558	152	128326	5.325	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.400%
Target Compounds						
					Qvalue	
13) Acetone	2.127	43	17525	4.873	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092024\
Data File : VV037440.D
Acq On : 20 Sep 2024 18:15
Operator : SY/MD
Sample : P4081-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AP7

Quant Time: Sep 20 23:39:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 20 23:32:27 2024
Response via : Initial Calibration

