

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092024\
 Data File : VV037458.D
 Acq On : 21 Sep 2024 02:31
 Operator : SY/MD
 Sample : P4107-09
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B71

Quant Time: Sep 21 04:55:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 21 01:15:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	290565	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	308546	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	139974	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	53354	3.130	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.600%
7) Chloroethane-d5	1.529	69	69909	3.933	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	2.053	65	29846	3.390	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.800%
20) 2-Butanone-d5	3.806	46	238002	46.717	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.440%
24) Chloroform-d	4.243	84	187646	4.534	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	4.941	65	91435	4.718	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	4.957	84	336607	4.069	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.400%
36) 1,2-Dichloropropane-d6	5.985	67	117545	4.569	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.400%
41) Toluene-d8	7.239	98	269058	3.610	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.200%
43) trans-1,3-Dichloroprop...	7.564	79	27884	3.295	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.800%
46) 2-Hexanone-d5	8.024	63	195734	47.341	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.680%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	101639	4.847	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.000%
66) 1,2-Dichlorobenzene-d4	11.558	152	118657	4.768	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092024\
Data File : VV037458.D
Acq On : 21 Sep 2024 02:31
Operator : SY/MD
Sample : P4107-09
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B71

Quant Time: Sep 21 04:55:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR091024WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 21 01:15:22 2024
Response via : Initial Calibration

