

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
 Data File : VV028009.D
 Acq On : 21 Sep 2022 12:08
 Operator : SY/MD
 Sample : VV0921WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK266

Quant Time: Sep 22 03:22:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	370548	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	353079	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	161380	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	140004	43.039	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.080%
7) Chloroethane-d5	1.561	69	122746	47.235	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.460%
11) 1,1-Dichloroethene-d2	2.098	63	202391	35.120	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.240%
21) 2-Butanone-d5	3.873	46	216387	111.865	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.860%
24) Chloroform-d	4.336	84	263529	48.252	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.500%
26) 1,2-Dichloroethane-d4	5.021	65	178733	50.498	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.000%
32) Benzene-d6	5.040	84	523986	51.489	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.980%
36) 1,2-Dichloropropane-d6	6.059	67	169821	54.040	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.080%
41) Toluene-d8	7.310	98	446494	47.911	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.820%
43) trans-1,3-Dichloroprop...	7.616	79	76499	48.680	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.360%
47) 2-Hexanone-d5	8.085	63	160001	103.932	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.930%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	230497	49.833	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.660%
66) 1,2-Dichlorobenzene-d4	11.619	152	171721	54.815	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028009.D
Acq On : 21 Sep 2022 12:08
Operator : SY/MD
Sample : VV0921WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK266

Quant Time: Sep 22 03:22:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration

