

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028013.D
 Acq On : 21 Sep 2022 13:51
 Operator : SY/MD
 Sample : N4620-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 22 03:23:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	359633	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	349866	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	155167	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	132406	41.938	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	83.880%	
7) Chloroethane-d5	1.558	69	112247	44.506	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	89.020%	
11) 1,1-Dichloroethene-d2	2.098	63	190698	34.095	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	68.200%	
21) 2-Butanone-d5	3.873	46	225111	119.907	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	119.910%	
24) Chloroform-d	4.339	84	263825	49.772	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.540%	
26) 1,2-Dichloroethane-d4	5.024	65	169906	49.461	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	98.920%	
32) Benzene-d6	5.040	84	511420	50.715	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	101.440%	
36) 1,2-Dichloropropane-d6	6.063	67	167336	53.738	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	107.480%	
41) Toluene-d8	7.310	98	432568	46.843	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.680%	
43) trans-1,3-Dichloroprop...	7.616	79	73743	47.357	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	94.720%	
47) 2-Hexanone-d5	8.085	63	168126	110.212	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	110.210%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	232298	50.683	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	101.360%	
66) 1,2-Dichlorobenzene-d4	11.622	152	163145	54.163	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	108.320%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028013.D
Acq On : 21 Sep 2022 13:51
Operator : SY/MD
Sample : N4620-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 22 03:23:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration

