

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028016.D
 Acq On : 21 Sep 2022 15:06
 Operator : SY/MD
 Sample : N4689-01 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AY5

Quant Time: Sep 22 03:23:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	349790	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	326229	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	148041	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.298	65	119084	38.780	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	77.560%	
7) Chloroethane-d5	1.558	69	104153	42.459	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	84.920%	
11) 1,1-Dichloroethene-d2	2.098	63	170909	31.417	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	62.840%	
21) 2-Butanone-d5	3.870	46	194393	106.458	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	106.460%	
24) Chloroform-d	4.339	84	244787	47.480	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.960%	
26) 1,2-Dichloroethane-d4	5.024	65	153550	45.958	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.920%	
32) Benzene-d6	5.040	84	440402	46.837	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.680%	
36) 1,2-Dichloropropane-d6	6.063	67	159700	55.002	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	110.000%	
41) Toluene-d8	7.310	98	375679	43.630	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.260%	
43) trans-1,3-Dichloroprop...	7.619	79	65210	44.912	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	89.820%	
47) 2-Hexanone-d5	8.085	63	147588	103.759	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	103.760%	
56) 1,1,2,2-Tetrachloroeth...	10.211	84	199082	46.584	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	93.160%	
66) 1,2-Dichlorobenzene-d4	11.622	152	149694	52.089	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	104.180%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028016.D
Acq On : 21 Sep 2022 15:06
Operator : SY/MD
Sample : N4689-01 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AY5

Quant Time: Sep 22 03:23:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration

