

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028038.D
 Acq On : 22 Sep 2022 01:03
 Operator : SY/MD
 Sample : N4659-06DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E62DL

Quant Time: Sep 22 06:25:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 05:25:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	358099	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	350811	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	150906	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	101235	32.202	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	64.400%		
7) Chloroethane-d5	1.561	69	93327	37.162	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	74.320%		
11) 1,1-Dichloroethene-d2	2.098	63	152154	27.320	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	54.640%#		
21) 2-Butanone-d5	3.876	46	211832	113.317	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	113.320%		
24) Chloroform-d	4.339	84	229672	43.515	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.020%		
26) 1,2-Dichloroethane-d4	5.024	65	151302	44.234	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.460%		
32) Benzene-d6	5.043	84	418781	41.417	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	82.840%		
36) 1,2-Dichloropropane-d6	6.063	67	144570	46.302	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	92.600%		
41) Toluene-d8	7.310	98	334096	36.082	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	72.160%#		
43) trans-1,3-Dichloroprop...	7.616	79	60281	38.608	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	77.220%		
47) 2-Hexanone-d5	8.085	63	149414	97.682	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.680%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	212298	46.195	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	92.400%		
66) 1,2-Dichlorobenzene-d4	11.622	152	138553	47.297	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	94.600%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.108	96	7894	3.038	ug/L	# 1
20) cis-1,2-Dichloroethene	3.902	96	119208	43.223	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028038.D
Acq On : 22 Sep 2022 01:03
Operator : SY/MD
Sample : N4659-06DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E62DL

Quant Time: Sep 22 06:25:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 05:25:17 2022
Response via : Initial Calibration

