

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
 Data File : VV028056.D
 Acq On : 22 Sep 2022 11:00
 Operator : SY/MD
 Sample : N4659-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0ED1

Quant Time: Sep 22 12:56:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 05:25:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	348273	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	337558	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	149098	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	143438	46.914	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.820%
7) Chloroethane-d5	1.561	69	117150	47.965	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.920%
11) 1,1-Dichloroethene-d2	2.098	63	202380	37.364	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.720%
21) 2-Butanone-d5	3.883	46	226993	124.853	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	124.850%
24) Chloroform-d	4.339	84	250043	48.711	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.420%
26) 1,2-Dichloroethane-d4	5.024	65	171356	51.511	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.020%
32) Benzene-d6	5.043	84	515004	52.933	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.860%
36) 1,2-Dichloropropane-d6	6.063	67	160702	53.489	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.980%
41) Toluene-d8	7.310	98	443985	49.832	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.660%
43) trans-1,3-Dichloroprop...	7.619	79	74263	49.430	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.860%
47) 2-Hexanone-d5	8.085	63	160119	108.791	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.790%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	230516	52.129	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.260%
66) 1,2-Dichlorobenzene-d4	11.622	152	166027	57.363	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.720%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.908	95	65104	26.088	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028056.D
Acq On : 22 Sep 2022 11:00
Operator : SY/MD
Sample : N4659-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0ED1

Quant Time: Sep 22 12:56:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 05:25:17 2022
Response via : Initial Calibration

