




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028027.D
Acq On : 21 Sep 2022 20:04
Operator : SY/MD
Sample : N4677-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E91

Manual IntegrationsAPPROVED

Quant Time: Sep 22 03:26:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 09/22/2022
Supervised By :Mahesh Dadoda 09/22/2022



(20) cis-1,2-Dichloroethene (T)

3.902min (+ 0.003) 2.91 ug/L

response 7441

Ion	Exp%	Act%
96.00	100.00	100.00
61.00	120.60	114.50
68.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
 Data File : VV028027.D
 Acq On : 21 Sep 2022 20:04
 Operator : SY/MD
 Sample : N4677-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E91

Manual IntegrationsAPPROVED

Quant Time: Sep 22 03:26:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 09/22/2022
 Supervised By :Mahesh Dadoda 09/22/2022



(20) cis-1,2-Dichloroethene (T)

3.902min (+ 0.003) 4.74 ug/L m

response 12090

Ion	Exp%	Act%
96.00	100.00	100.00
61.00	120.60	114.50
68.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
 Data File : VV028027.D
 Acq On : 21 Sep 2022 20:04
 Operator : SY/MD
 Sample : N4677-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E91

Manual IntegrationsAPPROVED

Quant Time: Sep 22 03:26:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 09/22/2022
 Supervised By :Mahesh Dadoda 09/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	331458	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	317122	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	139861	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	101655	34.935	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	69.880%	
7) Chloroethane-d5	1.558	69	95660	41.153	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	82.300%	
11) 1,1-Dichloroethene-d2	2.098	63	159495	30.940	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	61.880%	
21) 2-Butanone-d5	3.876	46	186963	108.052	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	108.050%	
24) Chloroform-d	4.339	84	212904	43.580	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	87.160%	
26) 1,2-Dichloroethane-d4	5.024	65	147872	46.706	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	93.420%	
32) Benzene-d6	5.043	84	408838	44.729	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	89.460%	
36) 1,2-Dichloropropane-d6	6.063	67	140919	49.927	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	99.860%	
41) Toluene-d8	7.313	98	345778	41.310	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	82.620%	
43) trans-1,3-Dichloroprop...	7.619	79	62921	44.580	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	89.160%	
47) 2-Hexanone-d5	8.085	63	148456	107.366	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	107.370%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	208465	50.180	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	100.360%	
66) 1,2-Dichlorobenzene-d4	11.622	152	138498	51.012	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	102.020%	
Target Compounds						
12) 1,1-Dichloroethene	2.111	96	13143	5.465	ug/L #	1
20) cis-1,2-Dichloroethene	3.902	96	12090m	4.736	ug/L	
30) 1,1,1-Trichloroethane	4.603	97	7755	1.986	ug/L #	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028027.D
Acq On : 21 Sep 2022 20:04
Operator : SY/MD
Sample : N4677-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E91

Manual IntegrationsAPPROVED

Quant Time: Sep 22 03:26:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 09/22/2022
Supervised By :Mahesh Dadoda 09/22/2022

