

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028014.D
 Acq On : 21 Sep 2022 14:16
 Operator : SY/MD
 Sample : N4619-19
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 22 03:23:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	343921	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	329296	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	150694	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	123702	40.971	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.940%
7) Chloroethane-d5	1.558	69	106740	44.256	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.520%
11) 1,1-Dichloroethene-d2	2.098	63	177118	33.114	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.220%
21) 2-Butanone-d5	3.870	46	221287	123.255	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	123.250%
24) Chloroform-d	4.339	84	236358	46.627	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.260%
26) 1,2-Dichloroethane-d4	5.024	65	170018	51.755	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.520%
32) Benzene-d6	5.043	84	481951	50.778	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.560%
36) 1,2-Dichloropropane-d6	6.062	67	157409	53.708	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.420%
41) Toluene-d8	7.310	98	409987	47.171	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.340%
43) trans-1,3-Dichloroprop...	7.619	79	71505	48.789	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.580%
47) 2-Hexanone-d5	8.085	63	149682	104.251	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.250%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	225416	52.254	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.500%
66) 1,2-Dichlorobenzene-d4	11.622	152	155513	53.161	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028014.D
Acq On : 21 Sep 2022 14:16
Operator : SY/MD
Sample : N4619-19
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 22 03:23:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration

