

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028021.D
 Acq On : 21 Sep 2022 17:34
 Operator : SY/MD
 Sample : N4677-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E82

Quant Time: Sep 22 03:24:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	342184	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	321850	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	145056	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	108613	36.156	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.320%
7) Chloroethane-d5	1.558	69	95847	39.941	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.880%
11) 1,1-Dichloroethene-d2	2.098	63	157322	29.562	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.120%#
21) 2-Butanone-d5	3.870	46	221855	124.198	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	124.200%
24) Chloroform-d	4.339	84	237438	47.078	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.160%
26) 1,2-Dichloroethane-d4	5.024	65	158744	48.568	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.140%
32) Benzene-d6	5.044	84	445315	48.004	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.000%
36) 1,2-Dichloropropane-d6	6.063	67	152002	53.063	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.120%
41) Toluene-d8	7.314	98	363940	42.841	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.680%
43) trans-1,3-Dichloroprop...	7.619	79	65945	46.036	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.080%
47) 2-Hexanone-d5	8.085	63	157029	111.898	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	111.900%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	214635	50.906	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.820%
66) 1,2-Dichlorobenzene-d4	11.622	152	140575	49.923	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.840%
Target Compounds						
13) Acetone	2.163	43	5614	3.527	ug/L	Qvalue 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028021.D
Acq On : 21 Sep 2022 17:34
Operator : SY/MD
Sample : N4677-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E82

Quant Time: Sep 22 03:24:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration





#13
 Acetone
 Concen: 3.527 ug/L
 RT: 2.163 min Scan# 34
 Delta R.T. -0.013 min
 Lab File: VV028021.D
 Acq: 21 Sep 2022 17:34

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E82

Tgt Ion: 43 Resp: 5614
 Ion Ratio Lower Upper
 43 100
 58 35.9 0.0 66.2

