

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028023.D
 Acq On : 21 Sep 2022 18:24
 Operator : SY/MD
 Sample : N4677-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E85

Quant Time: Sep 22 03:25:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 03:03:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	328118	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	292825	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	142011	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	100902	35.029	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.060%
7) Chloroethane-d5	1.558	69	89975	39.101	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.200%
11) 1,1-Dichloroethene-d2	2.098	63	148922	29.183	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.360%#
21) 2-Butanone-d5	3.870	46	188101	109.816	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.820%
24) Chloroform-d	4.339	84	219867	45.463	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.920%
26) 1,2-Dichloroethane-d4	5.024	65	140871	44.948	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.900%
32) Benzene-d6	5.043	84	394679	46.763	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.520%
36) 1,2-Dichloropropane-d6	6.063	67	139212	53.415	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.820%
41) Toluene-d8	7.310	98	340817	44.096	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.200%
43) trans-1,3-Dichloroprop...	7.619	79	60167	46.166	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.340%
47) 2-Hexanone-d5	8.085	63	145000	113.568	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.570%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	205137	53.476	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.960%
66) 1,2-Dichlorobenzene-d4	11.622	152	136637	49.565	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.120%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028023.D
Acq On : 21 Sep 2022 18:24
Operator : SY/MD
Sample : N4677-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E85

Quant Time: Sep 22 03:25:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 03:03:58 2022
Response via : Initial Calibration

