

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028032.D
 Acq On : 21 Sep 2022 22:34
 Operator : SY/MD
 Sample : N4689-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 22 06:24:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 05:25:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	365383	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	348731	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	155082	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	123550	38.517	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.040%
7) Chloroethane-d5	1.558	69	113173	44.167	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.340%
11) 1,1-Dichloroethene-d2	2.098	63	182268	32.075	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.160%
21) 2-Butanone-d5	3.876	46	231256	121.241	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	121.240%
24) Chloroform-d	4.343	84	254735	47.301	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.024	65	170646	48.895	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.800%
32) Benzene-d6	5.043	84	488982	48.648	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.300%
36) 1,2-Dichloropropane-d6	6.063	67	154935	49.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.840%
41) Toluene-d8	7.310	98	410215	44.566	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.140%
43) trans-1,3-Dichloroprop...	7.616	79	69607	44.847	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.700%
47) 2-Hexanone-d5	8.085	63	166854	109.734	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.730%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	211541	46.305	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	160583	53.341	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028032.D
Acq On : 21 Sep 2022 22:34
Operator : SY/MD
Sample : N4689-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 22 06:24:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 05:25:17 2022
Response via : Initial Calibration

