

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092122\
 Data File : VW028045.D
 Acq On : 22 Sep 2022 03:56
 Operator : SY/MD
 Sample : N4659-13
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0E80

Quant Time: Sep 22 06:26:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 22 05:25:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	337372	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	326764	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	145501	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	150543	50.829	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.660%
7) Chloroethane-d5	1.561	69	119990	50.715	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.440%
11) 1,1-Dichloroethene-d2	2.098	63	203683	38.820	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.640%
21) 2-Butanone-d5	3.883	46	207656	117.907	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.910%
24) Chloroform-d	4.339	84	253022	50.884	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.760%
26) 1,2-Dichloroethane-d4	5.024	65	170057	52.772	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.540%
32) Benzene-d6	5.043	84	506924	53.823	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.640%
36) 1,2-Dichloropropane-d6	6.063	67	166254	57.165	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	114.340%
41) Toluene-d8	7.310	98	442413	51.296	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.600%
43) trans-1,3-Dichloroprop...	7.619	79	69697	47.924	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.840%
47) 2-Hexanone-d5	8.085	63	160020	112.315	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.310%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	234349	54.746	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	109.500%
66) 1,2-Dichlorobenzene-d4	11.622	152	167660	59.359	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	118.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092122\
Data File : VV028045.D
Acq On : 22 Sep 2022 03:56
Operator : SY/MD
Sample : N4659-13
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0E80

Quant Time: Sep 22 06:26:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092022WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 22 05:25:17 2022
Response via : Initial Calibration

