

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092324\
 Data File : VV037487.D
 Acq On : 23 Sep 2024 16:59
 Operator : SY/MD
 Sample : VIBLK386
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK386

Quant Time: Sep 24 04:01:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Sep 24 03:56:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	807298	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	749398	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	374505	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	283637	38.408	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.820%
7) Chloroethane-d5	1.532	69	237432	41.600	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.200%
11) 1,1-Dichloroethene-d2	2.060	65	125930	40.572	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.140%
21) 2-Butanone-d5	3.793	46	260023	90.924	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.920%
24) Chloroform-d	4.243	84	550577	47.636	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.280%
26) 1,2-Dichloroethane-d4	4.937	65	344704	47.359	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.720%
32) Benzene-d6	4.953	84	1116587	49.331	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.660%
36) 1,2-Dichloropropane-d6	5.982	67	353260	49.829	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.660%
41) Toluene-d8	7.236	98	982257	48.084	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.160%
43) trans-1,3-Dichloroprop...	7.548	79	158253	48.249	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.500%
47) 2-Hexanone-d5	8.024	63	191779	99.118	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.120%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	403646	46.945	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.900%
66) 1,2-Dichlorobenzene-d4	11.554	152	374018	49.865	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092324\
Data File : VV037487.D
Acq On : 23 Sep 2024 16:59
Operator : SY/MD
Sample : VIBLK386
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK386

Quant Time: Sep 24 04:01:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM082624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 24 03:56:17 2024
Response via : Initial Calibration

