

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV092519\
 Data File : VV012975.D
 Acq On : 25 Sep 2019 14:30
 Operator : SY/MD
 Sample : K4958-11
 Misc : 4.72G/10ML/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ESNZ3

Quant Time: Sep 26 03:05:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM092319S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 26 01:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	745769	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	719862	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	275895	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	161264	13.94	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	55.76%
7) Chloroethane-d5	1.58	69	167282	17.88	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.52%
10) 1,1-Dichloroethene-d2	2.13	63	227040	11.80	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	47.20%
20) 2-Butanone-d5	3.96	46	99708	24.96	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	49.92%
24) Chloroform-d	4.40	84	391567	19.11	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.44%
26) 1,2-Dichloroethane-d4	5.08	65	231409	19.43	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	77.72%
29) Benzene-d6	5.10	84	764313	19.02	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.08%
33) 1,2-Dichloropropane-d6	6.11	67	255370	20.27	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.08%
37) Toluene-d8	7.36	98	626727	17.18	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.72%
38) trans-1,3-Dichloropropene-	7.66	79	85534	16.03	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	64.12%
39) 2-Hexanone-d5	8.14	63	64481	25.43	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	50.86%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	185614	15.51	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	62.04%
61) 1,2-Dichlorobenzene-d4	11.67	152	201984	18.84	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	75.36%

Target Compounds					Ovalue
13) Acetone	2.22	43	9144	2.153 ug/L	66
16) Methylene chloride	2.53	84	25267	2.267 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV092519\
Data File : VV012975.D
Acq On : 25 Sep 2019 14:30
Operator : SY/MD
Sample : K4958-11
Misc : 4.72G/10ML/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESNZ3

Quant Time: Sep 26 03:05:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM092319S.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 01:10:53 2019
Response via : Initial Calibration

