

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV092519\
Quantitation Report (Qedit)




Data File : VV012974.D
Acq On : 25 Sep 2019 14:03
Operator : SY/MD
Sample : K4958-09RE
Misc : 3.91G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESNZ1RE

Manual Integrations
APPROVED

MMDadoda
9/26/2019 3:31:58 PM

Quant Time: Sep 26 01:14:14 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM092319S.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 01:10:53 2019
Response via : Initial Calibration



TIC: VV012974.D

(21) 2-Butanone

4.013min (-0.016) 5.94ug/L

response 24492

Ion	Exp%	Act%
43.00	100	100
72.00	23.90	7.72#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV092519\
Quantitation Report (Qedit)




Data File : VV012974.D
Acq On : 25 Sep 2019 14:03
Operator : SY/MD
Sample : K4958-09RE
Misc : 3.91G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESNZ1RE

Manual Integrations
APPROVED

MMDadoda
9/26/2019 3:31:58 PM

Quant Time: Sep 26 01:14:14 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM092319S.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 01:10:53 2019
Response via : Initial Calibration



TIC: VV012974.D

(21) 2-Butanone

4.013min (-0.016) 7.07ug/L m

response 29130

Ion	Exp%	Act%
43.00	100	100
72.00	23.90	6.50#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV092519\
Quantitation Report (QT Reviewed)

Data File : VV012974.D
Acq On : 25 Sep 2019 14:03
Operator : SY/MD
Sample : K4958-09RE
Misc : 3.91G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESNZ1RE

Manual Integrations
APPROVED

MMDadoda
9/26/2019 3:31:58 PM

Quant Time: Sep 26 02:35:50 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM092319S.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 01:10:53 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	621878	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	536849	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	184895	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	177419	18.39	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.56%
7) Chloroethane-d5	1.58	69	166654	21.36	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.44%
10) 1,1-Dichloroethene-d2	2.13	63	223855	13.95	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.80%
20) 2-Butanone-d5	3.93	46	112942	33.90	ug/L	-0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	=	67.80%
24) Chloroform-d	4.40	84	341382	19.98	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.92%
26) 1,2-Dichloroethane-d4	5.08	65	216931	21.85	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.40%
29) Benzene-d6	5.10	84	662372	22.10	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.40%
33) 1,2-Dichloropropane-d6	6.11	67	224850	23.93	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.72%
37) Toluene-d8	7.36	98	491559	18.07	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.28%
38) trans-1,3-Dichloropropene-	7.66	79	68134	17.13	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.52%
39) 2-Hexanone-d5	8.13	63	72742	38.48	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.96%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	184135	20.63	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	82.52%
61) 1,2-Dichlorobenzene-d4	11.67	152	127065	17.69	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	70.76%#

Target Compounds

					Qvalue
13) Acetone	2.18	43	66167	18.686	ug/L 98
16) Methylene chloride	2.54	84	19997	2.151	ug/L 93
21) 2-Butanone	4.01	43	29130m	7.069	ug/L

09/26/2019

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV092519\
Quantitation Report (QT Reviewed)

Data File : VV012974.D
Acq On : 25 Sep 2019 14:03
Operator : SY/MD
Sample : K4958-09RE
Misc : 3.91G/10ML/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ESNZ1RE

Manual Integrations
APPROVED

MMDadoda
9/26/2019 3:31:58 PM

Quant Time: Sep 26 02:35:50 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM092319S.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 26 01:10:53 2019
Response via : Initial Calibration

