

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092622\
 Data File : VW028119.D
 Acq On : 26 Sep 2022 16:18
 Operator : SY/MD
 Sample : N4820-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CB8L2

Quant Time: Sep 26 21:31:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Sep 26 21:28:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	175567	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	175648	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	84994	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.297	65	71502	3.928	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.600%
7) Chloroethane-d5	1.561	69	66269	4.579	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	2.098	63	96981	3.072	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.400%
20) 2-Butanone-d5	3.921	46	123748	40.271	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.540%
24) Chloroform-d	4.339	84	135977	4.682	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
26) 1,2-Dichloroethane-d4	5.027	65	65729	4.756	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.200%
32) Benzene-d6	5.043	84	250111	4.343	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.066	67	79447	4.425	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.400%
41) Toluene-d8	7.310	98	192411	3.691	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.800%
43) trans-1,3-Dichloroprop...	7.619	79	27884	4.299	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.088	63	127135	46.967	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.940%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	60955	4.967	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.400%
66) 1,2-Dichlorobenzene-d4	11.619	152	84239	5.053	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092622\
Data File : VV028119.D
Acq On : 26 Sep 2022 16:18
Operator : SY/MD
Sample : N4820-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CB8L2

Quant Time: Sep 26 21:31:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Sep 26 21:28:54 2022
Response via : Initial Calibration

