

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092624\
 Data File : VV037546.D
 Acq On : 26 Sep 2024 14:40
 Operator : SY/MD
 Sample : P4131-22ME
 Misc : 5.27g/5mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4ER4ME

Quant Time: Sep 27 00:51:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 27 00:49:09 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	662347	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	627946	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	306567	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	259607	44.649	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.300%
7) Chloroethane-d5	1.506	69	110679	24.524	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	49.040%#
11) 1,1-Dichloroethene-d2	2.043	65	110557	45.512	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.020%
21) 2-Butanone-d5	3.802	46	205103	99.038	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.040%
24) Chloroform-d	4.236	84	448851	43.391	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.780%
26) 1,2-Dichloroethane-d4	4.934	65	272774	45.050	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.100%
32) Benzene-d6	4.947	84	909814	46.302	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.600%
36) 1,2-Dichloropropane-d6	5.982	67	286055	45.717	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.440%
41) Toluene-d8	7.236	98	798615	44.821	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.640%
43) trans-1,3-Dichloroprop...	7.548	79	126457	43.506	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.020%
47) 2-Hexanone-d5	8.027	63	127324	96.368	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.370%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	324131	44.583	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.160%
66) 1,2-Dichlorobenzene-d4	11.554	152	307580	47.613	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.220%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092624\
Data File : VV037546.D
Acq On : 26 Sep 2024 14:40
Operator : SY/MD
Sample : P4131-22ME
Misc : 5.27g/5mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4ER4ME

Quant Time: Sep 27 00:51:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 27 00:49:09 2024
Response via : Initial Calibration

