

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092624\
 Data File : VV037568.D
 Acq On : 26 Sep 2024 22:58
 Operator : SY/MD
 Sample : VV0926WBL02
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK215

Quant Time: Sep 27 02:23:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 27 02:22:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	676155	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	644466	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	305697	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	222530	37.491	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.980%
7) Chloroethane-d5	1.532	69	192126	41.701	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.400%
11) 1,1-Dichloroethene-d2	2.060	65	103316	41.663	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.320%
21) 2-Butanone-d5	3.793	46	220331	104.218	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.220%
24) Chloroform-d	4.243	84	477784	45.244	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.480%
26) 1,2-Dichloroethane-d4	4.937	65	294523	47.649	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.300%
32) Benzene-d6	4.953	84	947241	46.971	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.940%
36) 1,2-Dichloropropane-d6	5.982	67	308034	47.968	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.940%
41) Toluene-d8	7.236	98	845119	46.216	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.440%
43) trans-1,3-Dichloroprop...	7.548	79	138085	46.289	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.580%
47) 2-Hexanone-d5	8.024	63	134675	99.319	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.320%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	367001	49.186	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.380%
66) 1,2-Dichlorobenzene-d4	11.554	152	339457	52.697	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV092624\
Data File : VV037568.D
Acq On : 26 Sep 2024 22:58
Operator : SY/MD
Sample : VV0926WBL02
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK215

Quant Time: Sep 27 02:23:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 27 02:22:28 2024
Response via : Initial Calibration

