

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092922\
 Data File : VV028186.D
 Acq On : 29 Sep 2022 14:58
 Operator : SY/MD
 Sample : N4866-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXFQ0

Quant Time: Sep 30 01:47:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 29 01:25:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	162213	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	162455	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	76138	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	48274	2.871	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	57.400%
7) Chloroethane-d5	1.561	69	48096	3.597	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.000%
11) 1,1-Dichloroethene-d2	2.101	63	68585	2.351	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	47.000%#
20) 2-Butanone-d5	3.892	46	149144	52.531	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.060%
24) Chloroform-d	4.342	84	119739	4.462	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	5.027	65	60129	4.709	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%
32) Benzene-d6	5.043	84	213577	4.010	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.200%
36) 1,2-Dichloropropane-d6	6.066	67	70839	4.266	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.400%
41) Toluene-d8	7.313	98	164299	3.408	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.200%#
43) trans-1,3-Dichloroprop...	7.622	79	21150	3.526	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	70.600%
46) 2-Hexanone-d5	8.088	63	125918	50.295	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.580%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	57204	5.040	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	70587	4.727	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.600%
Target Compounds						Qvalue
5) Vinyl chloride	1.304	62	51445	4.090	ug/L	99
18) trans-1,2-Dichloroethene	2.760	96	2083	0.212	ug/L #	82
22) cis-1,2-Dichloroethene	3.902	96	15571	1.209	ug/L	96
33) Benzene	5.101	78	5503	0.122	ug/L	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092922\
Data File : VV028186.D
Acq On : 29 Sep 2022 14:58
Operator : SY/MD
Sample : N4866-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXFQ0

Quant Time: Sep 30 01:47:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 29 01:25:35 2022
Response via : Initial Calibration









#5
 Vinyl chloride
 Concen: 4.090 ug/L
 RT: 1.304 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VV028186.D
 Acq: 29 Sep 2022 14:58




Instrument :
 MSVOA_V
 ClientSampleId :
 EXFQ0


Tgt Ion: 62 Resp: 51445
 Ion Ratio Lower Upper
 62 100
 64 33.7 23.0 42.8



#18
 trans-1,2-Dichloroethene
 Concen: 0.212 ug/L
 RT: 2.760 min Scan# 535
 Delta R.T. 0.009 min
 Lab File: VV028186.D
 Acq: 29 Sep 2022 14:58

Tgt Ion: 96 Resp: 2083
 Ion Ratio Lower Upper
 96 100
 61 121.7 96.2 178.6
 98 43.7 45.7 84.9#









#22
 cis-1,2-Dichloroethene
 Concen: 1.209 ug/L
 RT: 3.902 min Scan# 890
 Delta R.T. 0.003 min
 Lab File: VV028186.D
 Acq: 29 Sep 2022 14:58

Instrument :
 MSVOA_V
 ClientSampleId :
 EXFQ0

Tgt Ion: 96 Resp: 15571
 Ion Ratio Lower Upper
 96 100
 61 120.3 87.2 161.9
 68 0.0 0.0 0.0



#33
 Benzene
 Concen: 0.122 ug/L
 RT: 5.101 min Scan# 1263
 Delta R.T. 0.013 min
 Lab File: VV028186.D
 Acq: 29 Sep 2022 14:58

Tgt Ion: 78 Resp: 5503

