

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
 Data File : VV037669.D
 Acq On : 30 Sep 2024 20:38
 Operator : SY/MD
 Sample : VV0930MBL02
 Misc : 5.00g/5mL/100ul/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK222

Quant Time: Oct 01 04:35:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 04:32:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	680651	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	644538	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	302893	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	258149	43.205	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.400%
7) Chloroethane-d5	1.532	69	207824	44.810	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.620%
11) 1,1-Dichloroethene-d2	2.060	65	108994	43.662	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.320%
21) 2-Butanone-d5	3.793	46	188258	88.459	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.460%
24) Chloroform-d	4.243	84	463340	43.587	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.180%
26) 1,2-Dichloroethane-d4	4.937	65	286808	46.094	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.180%
32) Benzene-d6	4.953	84	927614	45.992	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.980%
36) 1,2-Dichloropropane-d6	5.986	67	295036	45.939	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.880%
41) Toluene-d8	7.236	98	809182	44.245	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.500%
43) trans-1,3-Dichloroprop...	7.548	79	127089	42.598	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.200%
47) 2-Hexanone-d5	8.024	63	120508	88.861	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.860%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	316981	42.478	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.960%
66) 1,2-Dichlorobenzene-d4	11.554	152	301982	47.314	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
Data File : VV037669.D
Acq On : 30 Sep 2024 20:38
Operator : SY/MD
Sample : VV0930MBL02
Misc : 5.00g/5mL/100ul/5ml/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK222

Quant Time: Oct 01 04:35:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 04:32:07 2024
Response via : Initial Calibration

