

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
 Data File : VV037656.D
 Acq On : 30 Sep 2024 15:37
 Operator : SY/MD
 Sample : P4183-15ME
 Misc : 4.44g/5mL/100ul/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4EL8ME

Quant Time: Oct 01 02:09:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 02:04:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	718697	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	696900	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	359042	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	240715	38.154	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.300%
7) Chloroethane-d5	1.507	69	126054	25.740	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	51.480%#
11) 1,1-Dichloroethene-d2	2.047	65	117754	44.674	ug/L	-0.01
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.340%
21) 2-Butanone-d5	3.799	46	249604	111.076	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.080%
24) Chloroform-d	4.240	84	588186	52.402	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.800%
26) 1,2-Dichloroethane-d4	4.934	65	370544	56.400	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	112.800%
32) Benzene-d6	4.947	84	1131293	51.877	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.760%
36) 1,2-Dichloropropane-d6	5.982	67	372390	53.627	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.260%
41) Toluene-d8	7.236	98	1025721	51.871	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.740%
43) trans-1,3-Dichloroprop...	7.548	79	177337	54.974	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	109.940%
47) 2-Hexanone-d5	8.027	63	192012	130.949	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	130.950%#
56) 1,1,2,2-Tetrachloroeth...	10.146	84	462326	57.300	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	114.600%
66) 1,2-Dichlorobenzene-d4	11.554	152	451504	59.677	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	119.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
Data File : VV037656.D
Acq On : 30 Sep 2024 15:37
Operator : SY/MD
Sample : P4183-15ME
Misc : 4.44g/5mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4EL8ME

Quant Time: Oct 01 02:09:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 02:04:27 2024
Response via : Initial Calibration

