

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
 Data File : VV037666.D
 Acq On : 30 Sep 2024 19:12
 Operator : SY/MD
 Sample : P4162-15ME
 Misc : 5.20g/5mL/100ul/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4EJ0ME

Quant Time: Oct 01 02:11:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 02:04:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	875257	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	831883	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	424227	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	186374	24.257	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	48.520%#
7) Chloroethane-d5	1.510	69	104484	17.519	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	35.040%#
11) 1,1-Dichloroethene-d2	2.043	65	92642	28.860	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.720%#
21) 2-Butanone-d5	3.802	46	152308	55.655	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	55.650%
24) Chloroform-d	4.239	84	465145	34.028	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	68.060%#
26) 1,2-Dichloroethane-d4	4.934	65	297758	37.214	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.420%
32) Benzene-d6	4.947	84	915229	35.159	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.320%
36) 1,2-Dichloropropane-d6	5.982	67	306392	36.963	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	73.920%
41) Toluene-d8	7.236	98	834141	35.338	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	70.680%#
43) trans-1,3-Dichloroprop...	7.548	79	135554	35.203	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.400%
47) 2-Hexanone-d5	8.037	63	144703	82.672	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.670%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	357360	37.104	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	74.200%
66) 1,2-Dichlorobenzene-d4	11.554	152	358736	40.130	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.260%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
Data File : VV037666.D
Acq On : 30 Sep 2024 19:12
Operator : SY/MD
Sample : P4162-15ME
Misc : 5.20g/5mL/100ul/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4EJ0ME

Quant Time: Oct 01 02:11:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 02:04:27 2024
Response via : Initial Calibration

