

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
 Data File : VV037671.D
 Acq On : 30 Sep 2024 21:21
 Operator : SY/MD
 Sample : P4185-16DL 40X
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EY064DL

Quant Time: Oct 01 07:46:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 04:32:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	718644	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	693729	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	324554	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	187271	29.685	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	59.380%#
7) Chloroethane-d5	1.529	69	175566	35.853	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.700%
11) 1,1-Dichloroethene-d2	2.060	65	92011	34.910	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.820%
21) 2-Butanone-d5	3.793	46	203765	90.684	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.680%
24) Chloroform-d	4.243	84	467251	41.631	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.260%
26) 1,2-Dichloroethane-d4	4.934	65	289354	44.045	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.100%
32) Benzene-d6	4.953	84	903233	41.608	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.220%
36) 1,2-Dichloropropane-d6	5.982	67	299694	43.355	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.720%
41) Toluene-d8	7.236	98	798149	40.547	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.100%
43) trans-1,3-Dichloroprop...	7.548	79	130926	40.772	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.540%
47) 2-Hexanone-d5	8.021	63	125139	85.733	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.730%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	350750	43.670	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.340%
66) 1,2-Dichlorobenzene-d4	11.554	152	333640	48.785	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
Data File : VV037671.D
Acq On : 30 Sep 2024 21:21
Operator : SY/MD
Sample : P4185-16DL 40X
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EY064DL

Quant Time: Oct 01 07:46:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 04:32:07 2024
Response via : Initial Calibration

