

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
 Data File : VV037676.D
 Acq On : 30 Sep 2024 23:08
 Operator : SY/MD
 Sample : P4228-01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0CP6

Quant Time: Oct 01 07:47:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 04:32:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	659036	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	622259	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	297641	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	186015	32.153	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.300%
7) Chloroethane-d5	1.532	69	175397	39.059	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.120%
11) 1,1-Dichloroethene-d2	2.060	65	88735	36.712	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.420%
21) 2-Butanone-d5	3.793	46	192439	93.389	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.390%
24) Chloroform-d	4.243	84	476682	46.313	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.620%
26) 1,2-Dichloroethane-d4	4.937	65	290421	48.206	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.420%
32) Benzene-d6	4.954	84	902680	46.359	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.720%
36) 1,2-Dichloropropane-d6	5.986	67	297085	47.914	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.820%
41) Toluene-d8	7.240	98	795145	45.034	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.060%
43) trans-1,3-Dichloroprop...	7.548	79	130717	45.383	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.760%
47) 2-Hexanone-d5	8.024	63	120738	92.218	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.220%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	346119	48.043	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.080%
66) 1,2-Dichlorobenzene-d4	11.555	152	328541	52.383	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV093024\
Data File : VV037676.D
Acq On : 30 Sep 2024 23:08
Operator : SY/MD
Sample : P4228-01
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0CP6

Quant Time: Oct 01 07:47:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 04:32:07 2024
Response via : Initial Calibration

