

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100124\
 Data File : VV037694.D
 Acq On : 01 Oct 2024 11:54
 Operator : SY/MD
 Sample : P4228-03
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Oct 02 04:38:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 04:32:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	625655	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	606377	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	280807	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	245527	44.704	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	89.400%		
7) Chloroethane-d5	1.532	69	200863	47.116	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	94.240%		
11) 1,1-Dichloroethene-d2	2.060	65	107064	46.659	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	93.320%		
21) 2-Butanone-d5	3.796	46	171289	87.561	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	87.560%		
24) Chloroform-d	4.246	84	448316	45.881	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.760%		
26) 1,2-Dichloroethane-d4	4.937	65	270803	47.348	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.700%		
32) Benzene-d6	4.953	84	906378	47.768	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.540%		
36) 1,2-Dichloropropane-d6	5.985	67	284128	47.025	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	94.040%		
41) Toluene-d8	7.236	98	795823	46.253	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	92.500%		
43) trans-1,3-Dichloroprop...	7.551	79	123763	44.094	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	88.180%		
47) 2-Hexanone-d5	8.024	63	111309	87.243	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	87.240%		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	310878	44.281	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	88.560%		
66) 1,2-Dichlorobenzene-d4	11.554	152	296875	50.172	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	100.340%		
Target Compounds						Qvalue
10) 1,1,2-Trichloro-1,2,2-...	2.060	101	3671	0.823	ug/L	# 24

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100124\
Data File : VV037694.D
Acq On : 01 Oct 2024 11:54
Operator : SY/MD
Sample : P4228-03
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Oct 02 04:38:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 04:32:07 2024
Response via : Initial Calibration

