

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100124\
 Data File : VV037703.D
 Acq On : 01 Oct 2024 15:17
 Operator : SY/MD
 Sample : P4239-02
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A4FD2

Quant Time: Oct 02 04:39:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 04:32:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	621013	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	604312	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	276140	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	172527	31.648	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	63.300%		
7) Chloroethane-d5	1.532	69	163741	38.696	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	77.400%		
11) 1,1-Dichloroethene-d2	2.060	65	85362	37.479	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.960%		
21) 2-Butanone-d5	3.793	46	206228	106.209	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	106.210%		
24) Chloroform-d	4.243	84	449214	46.316	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.640%		
26) 1,2-Dichloroethane-d4	4.937	65	281848	49.647	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	99.300%		
32) Benzene-d6	4.953	84	859848	45.470	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	90.940%		
36) 1,2-Dichloropropane-d6	5.985	67	287893	47.811	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	95.620%		
41) Toluene-d8	7.239	98	757352	44.168	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.340%		
43) trans-1,3-Dichloroprop...	7.548	79	128144	45.811	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	91.620%		
47) 2-Hexanone-d5	8.024	63	122504	96.346	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	96.350%		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	350400	50.082	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	100.160%		
66) 1,2-Dichlorobenzene-d4	11.554	152	317752	54.608	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	109.220%		
Target Compounds					Qvalue	
13) Acetone	2.114	43	25540	13.418	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100124\
Data File : VV037703.D
Acq On : 01 Oct 2024 15:17
Operator : SY/MD
Sample : P4239-02
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A4FD2

Quant Time: Oct 02 04:39:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 04:32:07 2024
Response via : Initial Calibration

