

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100124\
 Data File : VV037709.D
 Acq On : 01 Oct 2024 17:27
 Operator : SY/MD
 Sample : P4237-22
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK002

Quant Time: Oct 02 04:40:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 01 04:32:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	607357	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	600747	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	272964	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	212881	39.928	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.860%
7) Chloroethane-d5	1.532	69	178330	43.091	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.180%
11) 1,1-Dichloroethene-d2	2.060	65	93841	42.128	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.260%
21) 2-Butanone-d5	3.796	46	166256	87.548	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.550%
24) Chloroform-d	4.243	84	412270	43.463	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.920%
26) 1,2-Dichloroethane-d4	4.937	65	252164	45.417	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.840%
32) Benzene-d6	4.953	84	809217	43.047	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.100%
36) 1,2-Dichloropropane-d6	5.982	67	263787	44.067	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.140%
41) Toluene-d8	7.239	98	700775	41.111	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.220%
43) trans-1,3-Dichloroprop...	7.548	79	111204	39.991	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.980%
47) 2-Hexanone-d5	8.024	63	110203	87.186	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.190%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	300579	43.216	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.440%
66) 1,2-Dichlorobenzene-d4	11.554	152	270493	47.027	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.060%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100124\
Data File : VV037709.D
Acq On : 01 Oct 2024 17:27
Operator : SY/MD
Sample : P4237-22
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Quant Time: Oct 02 04:40:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 01 04:32:07 2024
Response via : Initial Calibration

