

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100322\
 Data File : VW028283.D
 Acq On : 03 Oct 2022 17:10
 Operator : SY/MD
 Sample : N4870-09DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5H46DL

Quant Time: Oct 04 02:45:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 04 02:28:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	194850	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	195711	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	89904	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	70484	4.814	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	96.200%	
7) Chloroethane-d5	1.561	69	63540	5.127	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.600%	
11) 1,1-Dichloroethene-d2	2.102	63	101040	3.662	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.200%	
20) 2-Butanone-d5	3.899	46	146239	59.035	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.080%	
24) Chloroform-d	4.343	84	139830	5.272	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.400%	
26) 1,2-Dichloroethane-d4	5.027	65	65254	5.307	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.200%	
32) Benzene-d6	5.043	84	280911	5.232	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	104.600%	
36) 1,2-Dichloropropane-d6	6.066	67	83864	5.445	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	109.000%	
41) Toluene-d8	7.310	98	230100	4.660	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
43) trans-1,3-Dichloroprop...	7.622	79	22941	3.911	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.200%	
46) 2-Hexanone-d5	8.085	63	119925	49.951	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.900%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	59633	5.108	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	81960	5.297	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.000%	
Target Compounds						
						Qvalue
16) Methylene chloride	2.500	84	7680	0.488	ug/L	89
34) Trichloroethene	5.918	95	3874	0.266	ug/L	90
47) Tetrachloroethene	7.973	164	57857	4.538	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100322\
Data File : VV028283.D
Acq On : 03 Oct 2022 17:10
Operator : SY/MD
Sample : N4870-09DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5H46DL

Quant Time: Oct 04 02:45:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 04 02:28:45 2022
Response via : Initial Calibration

