

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW100322\
 Data File : VW028311.D
 Acq On : 04 Oct 2022 04:30
 Operator : SY/MD
 Sample : N4866-10DL 200X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXFQ6DL

Quant Time: Oct 04 06:18:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 04 02:28:45 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.612	114	184832	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.850	117	190330	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.246	152	88318	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.301	65	71170	5.125	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.400%	
7) Chloroethane-d5		1.564	69	63068	5.364	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.200%	
11) 1,1-Dichloroethene-d2		2.101	63	103577	3.957	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%	
20) 2-Butanone-d5		3.905	46	151998	64.686	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	129.380%	
24) Chloroform-d		4.339	84	142443	5.662	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.200%	
26) 1,2-Dichloroethane-d4		5.027	65	66579	5.708	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.200%	
32) Benzene-d6		5.043	84	290987	5.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.400%	
36) 1,2-Dichloropropane-d6		6.066	67	87552	5.845	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.000%	
41) Toluene-d8		7.313	98	239326	4.984	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.600%	
43) trans-1,3-Dichloroprop...		7.622	79	27452	4.812	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.200%	
46) 2-Hexanone-d5		8.088	63	125268	53.651	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.300%	
56) 1,1,2,2-Tetrachloroeth...		10.213	84	63803	5.620	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.400%	
66) 1,2-Dichlorobenzene-d4		11.619	152	89662	5.899	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.000%	
Target Compounds							Qvalue
5) Vinyl chloride		1.307	62	143379	9.037	ug/L	99
16) Methylene chloride		2.503	84	9389	0.629	ug/L	91
18) trans-1,2-Dichloroethene		2.757	96	967	0.070	ug/L	86
22) cis-1,2-Dichloroethene		3.902	96	182194	13.101	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100322\
Data File : VV028311.D
Acq On : 04 Oct 2022 04:30
Operator : SY/MD
Sample : N4866-10DL 200X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXFQ6DL

Quant Time: Oct 04 06:18:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 04 02:28:45 2022
Response via : Initial Calibration

