

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100424\
 Data File : VV037744.D
 Acq On : 04 Oct 2024 15:49
 Operator : SY/MD
 Sample : VIBLK399
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK399

Quant Time: Oct 07 04:13:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Oct 07 04:10:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	609086	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	599754	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	278987	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	242942	45.437	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.880%
7) Chloroethane-d5	1.532	69	203142	48.947	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.900%
11) 1,1-Dichloroethene-d2	2.059	65	107158	47.970	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.940%
21) 2-Butanone-d5	3.792	46	189323	99.412	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.410%
24) Chloroform-d	4.243	84	459681	48.323	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.640%
26) 1,2-Dichloroethane-d4	4.937	65	278198	49.964	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.920%
32) Benzene-d6	4.953	84	912006	48.595	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	5.985	67	294543	49.287	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.580%
41) Toluene-d8	7.239	98	790647	46.460	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.920%
43) trans-1,3-Dichloroprop...	7.548	79	124999	45.026	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.060%
47) 2-Hexanone-d5	8.024	63	112961	89.516	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.520%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	333673	48.053	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.100%
66) 1,2-Dichlorobenzene-d4	11.554	152	305496	51.966	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100424\
Data File : VV037744.D
Acq On : 04 Oct 2024 15:49
Operator : SY/MD
Sample : VIBLK399
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK399

Quant Time: Oct 07 04:13:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Oct 07 04:10:13 2024
Response via : Initial Calibration

