

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100522\
 Data File : VV028361.D
 Acq On : 05 Oct 2022 12:21
 Operator : SY/MD
 Sample : N4899-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5H00

Quant Time: Oct 06 02:32:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 04 02:28:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	217335	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	218975	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	102826	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	53222	3.259	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.200%
7) Chloroethane-d5	1.565	69	52770	3.817	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.400%
11) 1,1-Dichloroethene-d2	2.102	63	80009	2.599	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.000%#
20) 2-Butanone-d5	3.921	46	124992	45.238	ug/L	0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	90.480%
24) Chloroform-d	4.339	84	114368	3.866	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
26) 1,2-Dichloroethane-d4	5.027	65	56635	4.129	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
32) Benzene-d6	5.040	84	241752	4.025	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.400%
36) 1,2-Dichloropropane-d6	6.063	67	75935	4.407	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.200%
41) Toluene-d8	7.310	98	194761	3.525	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.600%
43) trans-1,3-Dichloroprop...	7.622	79	22623	3.447	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.000%
46) 2-Hexanone-d5	8.088	63	102733	38.244	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.480%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	50751	3.885	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.800%
66) 1,2-Dichlorobenzene-d4	11.619	152	72942	4.122	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100522\
Data File : VV028361.D
Acq On : 05 Oct 2022 12:21
Operator : SY/MD
Sample : N4899-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5H00

Quant Time: Oct 06 02:32:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 04 02:28:45 2022
Response via : Initial Calibration

