

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100522\
 Data File : VV028373.D
 Acq On : 05 Oct 2022 17:02
 Operator : SY/MD
 Sample : N4872-01DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXFW7DL

Quant Time: Oct 06 02:33:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 04 02:28:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	186664	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	183721	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	88588	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	48266	3.441	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.800%
7) Chloroethane-d5	1.565	69	49100	4.135	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	2.105	63	74839	2.831	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.600%#
20) 2-Butanone-d5	3.921	46	122273	51.525	ug/L	0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.040%
24) Chloroform-d	4.343	84	126901	4.994	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.800%
26) 1,2-Dichloroethane-d4	5.031	65	53261	4.521	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.400%
32) Benzene-d6	5.047	84	233998	4.643	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	6.066	67	73772	5.103	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.000%
41) Toluene-d8	7.313	98	185935	4.011	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.200%
43) trans-1,3-Dichloroprop...	7.622	79	22079	4.010	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.200%
46) 2-Hexanone-d5	8.088	63	95933	42.565	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.140%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	50706	4.627	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.600%
66) 1,2-Dichlorobenzene-d4	11.619	152	75905	4.979	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.600%
Target Compounds						Qvalue
16) Methylene chloride	2.503	84	3999	0.265	ug/L	94
22) cis-1,2-Dichloroethene	3.918	96	1929	0.137	ug/L	97
25) Chloroform	4.371	83	92101	3.536	ug/L	97
31) Carbon tetrachloride	4.822	117	151922	7.473	ug/L	99
47) Tetrachloroethene	7.969	164	755	0.063	ug/L	80
51) Chlorobenzene	8.883	112	20480	0.535	ug/L	96
64) 1,3-Dichlorobenzene	11.178	146	103996	3.905	ug/L	97
65) 1,4-Dichlorobenzene	11.268	146	292415	10.967	ug/L	98
67) 1,2-Dichlorobenzene	11.638	146	297900	12.335	ug/L	98
70) 1,2,4-trichlorobenzene	13.259	180	7827	0.488	ug/L	97
72) 1,2,3-Trichlorobenzene	13.744	180	3564	0.246	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100522\
Data File : VV028373.D
Acq On : 05 Oct 2022 17:02
Operator : SY/MD
Sample : N4872-01DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EXFW7DL

Quant Time: Oct 06 02:33:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 04 02:28:45 2022
Response via : Initial Calibration

