

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100722\
 Data File : VV028418.D
 Acq On : 07 Oct 2022 19:36
 Operator : SY/MD
 Sample : N4893-18DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5J51DL

Quant Time: Oct 08 01:01:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 07 23:51:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	386226	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	344455	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	164378	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	116795	39.191	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	78.380%		
7) Chloroethane-d5	1.564	69	101503	42.955	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	85.900%		
11) 1,1-Dichloroethene-d2	2.105	63	157493	31.052	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.100%		
21) 2-Butanone-d5	3.902	46	268297	102.842	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	102.840%		
24) Chloroform-d	4.342	84	259136	47.411	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.820%		
26) 1,2-Dichloroethane-d4	5.027	65	178788	50.721	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.440%		
32) Benzene-d6	5.043	84	543583	49.249	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.500%		
36) 1,2-Dichloropropane-d6	6.066	67	181624	50.470	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	100.940%		
41) Toluene-d8	7.310	98	485711	49.151	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	98.300%		
43) trans-1,3-Dichloroprop...	7.619	79	87805	47.869	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.740%		
47) 2-Hexanone-d5	8.085	63	190133	97.125	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	97.120%		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	240953	48.930	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	97.860%		
66) 1,2-Dichlorobenzene-d4	11.615	152	173788	51.024	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	102.040%		
Target Compounds					Qvalue	
46) Tetrachloroethene	7.969	164	97571	48.317	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100722\
Data File : VV028418.D
Acq On : 07 Oct 2022 19:36
Operator : SY/MD
Sample : N4893-18DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5J51DL

Quant Time: Oct 08 01:01:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 07 23:51:49 2022
Response via : Initial Calibration

