

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100722\
 Data File : VV028428.D
 Acq On : 08 Oct 2022 00:01
 Operator : SY/MD
 Sample : VV1007WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK270

Quant Time: Oct 08 02:11:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 08 02:10:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	370317	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	336112	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	161402	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	122010	42.700	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.400%
7) Chloroethane-d5	1.564	69	101029	44.591	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.180%
11) 1,1-Dichloroethene-d2	2.105	63	162927	33.503	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.000%
21) 2-Butanone-d5	3.905	46	251299	100.465	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.470%
24) Chloroform-d	4.343	84	253375	48.349	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.700%
26) 1,2-Dichloroethane-d4	5.027	65	174507	51.633	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.260%
32) Benzene-d6	5.043	84	539837	50.123	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.240%
36) 1,2-Dichloropropane-d6	6.066	67	175021	49.842	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.680%
41) Toluene-d8	7.310	98	482357	50.023	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.040%
43) trans-1,3-Dichloroprop...	7.616	79	85401	47.714	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.420%
47) 2-Hexanone-d5	8.085	63	184047	96.349	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.350%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	233952	48.687	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.380%
66) 1,2-Dichlorobenzene-d4	11.616	152	174852	52.283	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100722\
Data File : VV028428.D
Acq On : 08 Oct 2022 00:01
Operator : SY/MD
Sample : VV1007WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK270

Quant Time: Oct 08 02:11:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 08 02:10:35 2022
Response via : Initial Calibration

