

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100722\
 Data File : VV028403.D
 Acq On : 07 Oct 2022 13:29
 Operator : SY/MD
 Sample : VV1007WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK269

Quant Time: Oct 07 23:52:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 07 23:51:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	468507	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	419424	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	202604	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	155183	42.927	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.860%
7) Chloroethane-d5	1.564	69	129832	45.294	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.580%
11) 1,1-Dichloroethene-d2	2.101	63	209999	34.132	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.260%
21) 2-Butanone-d5	3.902	46	317718	100.398	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.400%
24) Chloroform-d	4.342	84	300354	45.302	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	5.027	65	206184	48.220	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.440%
32) Benzene-d6	5.043	84	656918	48.879	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.760%
36) 1,2-Dichloropropane-d6	6.066	67	214532	48.958	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.920%
41) Toluene-d8	7.310	98	581651	48.339	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.680%
43) trans-1,3-Dichloroprop...	7.616	79	105949	47.437	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.880%
47) 2-Hexanone-d5	8.085	63	231965	97.314	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.310%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	286338	47.753	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.500%
66) 1,2-Dichlorobenzene-d4	11.615	152	206039	49.079	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100722\
Data File : VV028403.D
Acq On : 07 Oct 2022 13:29
Operator : SY/MD
Sample : VV1007WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK269

Quant Time: Oct 07 23:52:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 07 23:51:49 2022
Response via : Initial Calibration

