

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100724\
 Data File : VV037748.D
 Acq On : 07 Oct 2024 09:35
 Operator : SY/MD
 Sample : VV1007MBL01
 Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK232

Quant Time: Oct 08 04:30:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 08 04:30:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	631576	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	608983	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	283721	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	257821	46.503	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	93.000%		
7) Chloroethane-d5	1.532	69	214288	49.794	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	99.580%		
11) 1,1-Dichloroethene-d2	2.060	65	111085	47.957	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	95.920%		
21) 2-Butanone-d5	3.789	46	179882	91.091	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	91.090%		
24) Chloroform-d	4.240	84	487657	49.439	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	98.880%		
26) 1,2-Dichloroethane-d4	4.934	65	294726	51.047	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	102.100%		
32) Benzene-d6	4.950	84	963631	50.568	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	101.140%		
36) 1,2-Dichloropropane-d6	5.979	67	307810	50.726	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	101.460%		
41) Toluene-d8	7.236	98	842625	48.764	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.520%		
43) trans-1,3-Dichloroprop...	7.548	79	132684	47.070	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	94.140%		
47) 2-Hexanone-d5	8.021	63	102087	79.673	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	79.670%		
56) 1,1,2,2-Tetrachloroeth...	10.146	84	326234	46.270	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	92.540%		
66) 1,2-Dichlorobenzene-d4	11.554	152	318955	53.350	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	106.700%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100724\
Data File : VV037748.D
Acq On : 07 Oct 2024 09:35
Operator : SY/MD
Sample : VV1007MBL01
Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK232

Quant Time: Oct 08 04:30:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 08 04:30:01 2024
Response via : Initial Calibration

