

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100924\
 Data File : VV037779.D
 Acq On : 09 Oct 2024 14:44
 Operator : SY/MD
 Sample : P4349-07
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOA7

Quant Time: Oct 10 03:12:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 10 03:11:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	617604	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	601151	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	266329	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	211823	39.070	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.140%
7) Chloroethane-d5	1.532	69	181499	43.129	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.260%
11) 1,1-Dichloroethene-d2	2.059	65	96822	42.745	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.500%
21) 2-Butanone-d5	3.799	46	155323	80.434	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.430%
24) Chloroform-d	4.243	84	431094	44.693	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.380%
26) 1,2-Dichloroethane-d4	4.937	65	264143	46.785	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.580%
32) Benzene-d6	4.953	84	850703	45.223	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.440%
36) 1,2-Dichloropropane-d6	5.982	67	275866	46.054	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.100%
41) Toluene-d8	7.236	98	749219	43.923	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.840%
43) trans-1,3-Dichloroprop...	7.551	79	118314	42.519	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.040%
47) 2-Hexanone-d5	8.024	63	94189	74.466	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.470%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	300676	43.201	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.400%
66) 1,2-Dichlorobenzene-d4	11.554	152	289705	51.622	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.240%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV100924\
Data File : VV037779.D
Acq On : 09 Oct 2024 14:44
Operator : SY/MD
Sample : P4349-07
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOAH7

Quant Time: Oct 10 03:12:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 10 03:11:31 2024
Response via : Initial Calibration

