

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101024\
 Data File : VV037794.D
 Acq On : 10 Oct 2024 11:07
 Operator : SY/MD
 Sample : P4361-01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOAY2

Quant Time: Oct 11 05:11:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 11 05:10:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	585162	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	574312	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	258186	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	208819	40.652	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.300%
7) Chloroethane-d5	1.532	69	177852	44.605	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.220%
11) 1,1-Dichloroethene-d2	2.060	65	92793	43.238	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.480%
21) 2-Butanone-d5	3.799	46	161686	88.371	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.370%
24) Chloroform-d	4.243	84	429083	46.951	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.900%
26) 1,2-Dichloroethane-d4	4.937	65	262993	49.164	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.320%
32) Benzene-d6	4.953	84	851161	47.362	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.720%
36) 1,2-Dichloropropane-d6	5.985	67	272595	47.635	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.260%
41) Toluene-d8	7.236	98	746316	45.798	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.600%
43) trans-1,3-Dichloroprop...	7.548	79	117059	44.034	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.060%
47) 2-Hexanone-d5	8.024	63	100697	83.332	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.330%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	301189	45.297	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	11.551	152	280188	51.500	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101024\
Data File : VV037794.D
Acq On : 10 Oct 2024 11:07
Operator : SY/MD
Sample : P4361-01
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOAY2

Quant Time: Oct 11 05:11:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 11 05:10:27 2024
Response via : Initial Calibration

