

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101024\
 Data File : VV037805.D
 Acq On : 10 Oct 2024 15:04
 Operator : SY/MD
 Sample : P4379-06
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EOLG9

Quant Time: Oct 11 05:14:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 11 05:10:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	593711	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	596472	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	258659	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	198319	38.052	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.100%
7) Chloroethane-d5	1.532	69	169056	41.789	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.580%
11) 1,1-Dichloroethene-d2	2.056	65	92106	42.300	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.600%
21) 2-Butanone-d5	3.796	46	163034	87.825	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.820%
24) Chloroform-d	4.243	84	421014	45.405	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.800%
26) 1,2-Dichloroethane-d4	4.934	65	254258	46.847	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.700%
32) Benzene-d6	4.950	84	822758	44.081	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.160%
36) 1,2-Dichloropropane-d6	5.982	67	265705	44.706	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.420%
41) Toluene-d8	7.236	98	724401	42.801	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.600%
43) trans-1,3-Dichloroprop...	7.551	79	110887	40.163	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.320%
47) 2-Hexanone-d5	8.024	63	105385	83.972	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.970%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	308374	44.654	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.300%
66) 1,2-Dichlorobenzene-d4	11.551	152	280587	51.479	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101024\
Data File : VV037805.D
Acq On : 10 Oct 2024 15:04
Operator : SY/MD
Sample : P4379-06
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EOLG9

Quant Time: Oct 11 05:14:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 11 05:10:27 2024
Response via : Initial Calibration

